



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

AUG 18 2000

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Kontur Kontakt Lens Co., Inc.
c/o Mr. Kevin M. Randall
P.O. Box 253
Lafayette, CO 80026

Re: K001908
Trade Name: Kontur 55 Soft Lens (parametric release)
Regulatory Class: II
Product Code: 86 LPL
Dated: June 15, 2000
Received: June 22, 2000

Dear Mr. Randall:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 – Mr. Kevin M. Randall

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-6413. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in cursive script that reads "Nancy C. Brogdon".


Nancy C. Brogdon
Acting Director
Division of Ophthalmic and Ear,
Nose and Throat Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

**INDICATION FOR USE
FOR
KONTUR (methafilcon A) SOFT LENSES**

The spherical Kontur 55 (methafilcon A) Soft Lens is indicated for daily wear for the correction of visual acuity in aphakic and not-aphakic persons with nondiseased eyes that are myopic or hyperopic. The lens may be worn by persons who exhibit refractive astigmatism of 1.5 diopters (D) or less that does not interfere with visual acuity. The spherical lens ranges in powers from -20.00 D to +20.00 D. The toric Kontur 55 (methafilcon A) Soft Lens is indicated for daily wear for the correction of visual acuity in not-aphakic persons with nondiseased eyes that are myopic, hyperopic or have refractive astigmatism not exceeding 5.00 D that does not interfere with visual acuity. The toric lens ranges in powers from -20.00 D to +10.00 D.

**PLEASE DO NOT WRITE BELOW THIS LINE
CONTINUE ON ANOTHER PAGE IF NEEDED**

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Ophthalmic Devices
510(k) Number K001908



Prescription Use ☒

OR

Over-The-Counter Use _____

(Optional Format 1-2-96)